

Certificate of Analysis

Product Name	Human anti-mutated citrullinated vimentin (MCV) antibody ELISA Kit		
Catalog	OKCA00429		
Lot	KH0776		
Test parameter	specification	test result	approval
Positive control	>0.8	1.2457	approved
Negative control	<0.1	0.0891	approved
CV	<15%	11.49	approved
Positive rate	10, Positive	100%	approved
Negative rate	10, Negative	100%	approved