



Certificate of Analysis

Product: IgE EIA TEST KIT
Catalog Number: OKBA00017
Lot Number: KF1072
Expiration Date: 08-2022

Kit Component	Lot No.	Contents	Exp. Date
IgE Ab Coated Wells	KF1072	1 ea	08-2022
Standard Set	KF1072	1 set	08-2022
Conjugate Reagent	KF1072	18 mL	08-2022
TMB Reagent	KF1072	11 mL	08-2022
Stop Solution (1N HCl)	KF1072	11 mL	08-2022
Zero Buffer	KF1072	13 mL	08-2022
Bio-Rad Controls (not included)	---	---	---

Standard Curve

Concentration (IU/mL)	O.D. 450 nm
0	0.093
10	0.222
50	0.656
100	1.134
400	2.464
800	3.016

This standard Curve was obtained on the date of performance testing of the above components. It is for the purpose of illustration only. Each laboratory should establish its own standard curve.

Control Values

Control No.	Mean (IU/mL)	Range
1	293	143 – 422
2	55.0	40.0 – 65.9
3	74.3	47.8 – 83.6