



Certificate of Analysis

Product: TSH EIA TEST KIT
Catalog Number: OKBA00026
Lot Number: KF0074
Expiration Date: 07-2021

Kit Component	Lot No.	Contents	Exp. Date
TSH Ab Coated Wells	KF0074	1 ea	07-2021
Standard Set	KF0074	1 set	07-2021
Conjugate Reagent	KF0074	13 mL	07-2021
TMB Reagent	KF0074	11 mL	07-2021
Stop Solution	KF0074	11 mL	07-2021
Bio-Rad Controls (not included)	---	---	---

Standard Curve

Concentration (μ U/mL)	O.D. 450 nm
0	0.047
0.5	0.128
2.0	0.349
5.0	0.814
10.0	1.493
25.0	2.992

This Standard Curve was obtained on the date of performance testing of the above components. It is for the purpose of illustration only. Each laboratory should establish its own standard curve.

Control Values

Control No.	Mean (μ U/mL)	Range
1	0.661	0.144 – 0.867
2	7.05	4.02 – 7.83
3	> 25.0	> 25.0

AVIVA SYSTEMS BIOLOGY

7700 Ronson Rd, Suite 100 San Diego, CA 92111 USA | Tel: (858)552-6979 | techsupport@avivasysbio.com