

# Gp130/IL6ST ELISA Kit (Mouse) (OKBB00409)

# Instruction for Use

For the quantitative measurement of Mouse Gp130 in cell culture supernatants, serum and plasma (heparin, EDTA, citrate).

This product is intended for research use only.



## **Table of Contents**

1.	Background	2
	Assay Summary	
3.	Precautions	3
4.	Storage and Stability	3
5.	Kit Components	3
6.	Required Materials Not Supplied	3
7.	Technical Application Tips	4
8.	Reagent Preparation	4
9.	Sample Preparation	6
10.	Assay Procedure	7
11.	Calculation of Results	8
12.	Typical Expected Data	8
13.	Technical Resources	. 10



### 1. Background


#### **Principle**

Aviva Systems Biology Gp130/IL6ST ELISA Kit (Mouse) (OKBB00409) is based on standard sandwich enzymelinked immuno-sorbent assay technology. A mouse monoclonal antibody specific for Gp130 has been precoated onto 96-wellplate (12 x 8 Well Strips). Standards (NSO; Q23-E617) and test samples are added to the wells and incubated. After washing, a biotinylated polyclonal goat detector antibody specific for Gp130 is added, incubated and followed by washing. Avidin-Biotin-Peroxidase Complex is then added, incubated and unbound conjugate is washed away. An enzymatic reaction is visualized through the addition of TMB substrate which is catalyzed by HRP to produce a blue color product that changes yellow after adding acidic stop solution. The density of yellow coloration read by absorbance at 450 nm and is quantitatively proportional to the amount of sample Mouse Gp130 captured in well.

#### **Background**

Gp130 (Glycoprotein 130) also known as IL6ST or CD130, is a transmembrane protein which is the founding member of the class of all cytokine receptors. The GP130 gene is mapped on 5q11.2. Gp130 showed an apparent molecular mass of 130 kD by SDS-PAGE. Gp130 is involved in the formation of high-affinity IL6-binding sites and in IL6 signal transduction. Sgp130 has a role in modulating signals transduced by membrane-bound gp130. Gp130 is an important part of many different types of signaling complexes. Gp130 molecules with mutations in either of these segments could not transduce a growth signal.

## 2. Assay Summary





#### 3. Precautions

- Read instructions fully prior to beginning use of the assay kit.
- Any deviations or modifications from the described method or use of other reagents could result in a reduction of performance.
- Reduce exposure to potentially harmful substances by wearing personal protective lab equipment including lab coats, gloves and glasses.
- For information on hazardous substances included in the kit please refer to the Material Safety Data Sheet (MSDS).

## 4. Storage and Stability

• Upon receipt store kit at 4°C for 6 months or -20°C for 12 months. Avoid multiple freeze/thaw cycles.

## 5. Kit Components

• The following reagents are the provided contents of the kit.

Description	Quantity	
96-Well plate Pre-coated with Anti-Mouse Gp130 Antibody	1 (12 x 8 Well Strip)	
Lyophilized Recombinant Mouse Gp130 Standard	10 ng/tube × 2	
100X Biotinylated Anti-Mouse Gp130 Antibody	130 µL	
100X Avidin-Biotin-Peroxidase Complex (ABC)	130 µL	
Sample Diluent Buffer	30 mL	
Antibody Diluent Buffer	12 mL	
ABC Diluent Buffer	12 mL	
TMB Color Developing Agent	10 mL	
TMB Stop Solution	10 mL	
10X Wash Buffer	30 mL	

## 6. Required Materials Not Supplied

- Microplate reader capable of reading absorbance at 450 nm.
- Automated plate washer (optional).
- Pipettes capable of precisely dispensing 0.5 µL through 1 mL volumes of aqueous solutions.
- Pipettes or volumetric glassware capable of precisely measuring 1 mL through 100 mL of aqueous solutions.
- New, clean tubes and/or micro-centrifuge tubes for the preparation of standards or samples.
- Absorbent paper or paper toweling.
- Distilled or deionized ultrapure water.



## 7. Technical Application Tips

- Do not mix or substitute components from other kits.
- To ensure the validity of experimental operation, it is recommended that pilot experiments using standards and a small selection of sample dilutions to ensure optimal dilution range for quantitation.
- Samples exhibiting OD measurements higher than the highest standard should be diluted further in the appropriate sample dilution buffers.
- Inspect all reagents prior to use. Components should contain no particulates or cloudiness and should be colorless.
- Prior to using the kit, briefly spin component tubes to collect all reagent at the bottom.
- Replicate wells are recommended for standards and samples.
- Cover microplate while incubating to prevent evaporation.
- Do not allow the microplate wells dry at any point during the assay procedure.
- Do not reuse tips or tube to prevent cross contamination.
- Avoid causing bubbles or foaming when pipetting, mixing or reconstituting.
- Completely remove of all liquids when washing to prevent cross contamination.
- Prepare reagents immediately prior to use and do not store, with the exception of the top standard.
- Equilibrate all materials to ambient room temperature prior to use (standards exception).
- For optimal results in inter- intra-assay consistency, equilibrate all materials to 37°C prior to performing assay (standards exception) and perform all incubations at 37°C.
- Pipetting less than 1 µL is not recommended for optimal assay accuracy.
- Once the procedure has been started, all steps should be completed without interruption. Ensure that all reagents, materials and devices are ready at the appropriate time.
- Incubation times will affect results. All wells should be handled in the same sequential order and time intervals for optimal results.
- Samples containing precipitates or fibrin strands or which are hemolytic of lipemic might cause inaccurate results due to interfering factors.

## 8. Reagent Preparation

• Equilibrate all materials to room temperature prior to use and use prepare immediately prior to use.

#### 8.1 1X Biotinylated Anti-Mouse Gp130 Antibody

- 8.1.1 Prepare the **1X Biotinylated Anti-Mouse Gp130 Antibody** immediately prior to use by diluting the **100X Biotinylated Anti-Mouse Gp130 Antibody** 1:100 with Antibody Diluent Buffer.
- 8.1.2 For each well to be used in the experiment prepare 120 μL by adding1.2 μL of **100X Biotinylated Anti-Mouse Gp130 Antibody** to 118.8 μL Antibody Diluent Buffer.
- 8.1.3 Mix thoroughly and gently. Hold no longer than 2 hours prior to using in procedure.

#### 8.2 1X Avidin-Biotin-Peroxidase Complex (ABC)

- 8.2.1 Prepare the **1X Avidin-Biotin-Peroxidase Complex (ABC)** immediately prior to use by diluting the **100X Avidin-Biotin-Peroxidase Complex (ABC)** 1:100 with **ABC Dilution Buffer**.
- 8.2.2 For each well to be used in the experiment prepare 120  $\mu$ L, by adding 1.2  $\mu$ L of **100X Avidin-Biotin-Peroxidase Complex (ABC)** to 118.8  $\mu$ L **ABC Dilution Buffer**.
- 8.2.3 Mix thoroughly and gently. Hold no longer than 2 hours prior to using in procedure.



#### 8.3 Gp130 Assay Standards

- 8.3.1 Prepare the Gp130 standards no greater than 2 hours prior to performing experiment. Standards should be held on ice until use in the experiment.
- 8.3.2 Reconstitute one of the provided 10 ng Lyophilized Recombinant Mouse Gp130 Standard. Use one for each experiment. Prepare a stock 10,000 pg/mL Mouse Gp130 Standard by reconstituting one tube of Lyophilized Recombinant Mouse Gp130 Standard as follows:
  - 8.3.2.1 Gently spin or tap the vial to collect all material at the bottom.
  - 8.3.2.2 Add 1 mL of Sample Diluent Buffer to the vial.
  - 8.3.2.3 Seal then mix gently and thoroughly.
  - 8.3.2.4 Leave the vial to sit at ambient temperature for 10 minutes.
- 8.3.3 Prepare a set of seven serially diluted standards as follows:
  - 8.3.3.1 Label tubes with numbers 1 8.
  - 8.3.3.2 Add 300  $\mu$ L of **Sample Diluent Buffer** to Tube #'s 2 8.
  - 8.3.3.3 Prepare an **8,000 pg/mL Standard #1** by adding 800  $\mu$ L of the 10,000 pg/mL reconstituted **Mouse Gp130 Standard** to 200  $\mu$ L of **Sample Diluent Buffer** in Tube#1. Mix gently and thoroughly.
  - 8.3.3.4 Prepare **Standard #2** by adding 300 μL of **8,000 pg/mL Standard #1** from Tube #1 to Tube #2. Mix gently and thoroughly.
  - 8.3.3.5 Prepare **Standard #3** by adding 300  $\mu$ L of **Standard #2** from Tube #2 to Tube #3. Mix gently and thoroughly.
  - 8.3.3.6 Prepare further serial dilutions through Tube #7. Reference the table below as a guide for serial dilution scheme.
  - 8.3.3.7 Tube #8 is a blank standard (only **Sample Diluent Buffer**), which should be included with every experiment.

Standard Number (Tube)	Sample To Dilute	Volume Standard (µL)	Volume Sample Diluent Buffer (µL)	Total Volume (μL)	Final Concentration
Tube #1	10,000 pg/mL of Mouse Gp130 Standard	800	200	1,000	8,000 pg/mL
Tube #2	8,000 pg/mL	300	300	600	4,000 pg/mL
Tube #3	4,000 pg/mL	300	300	600	2,000 pg/mL
Tube #4	2,000 pg/mL	300	300	600	1,000 pg/mL
Tube #5	1,000 pg/mL	300	300	600	500 pg/mL
Tube #6	500 pg/mL	300	300	600	250 pg/mL
Tube #7	250 pg/mL	300	300	600	125 pg/mL
Tube #8	NA	0	300	300	0.0 (Blank)





#### 8.4 1X Wash Buffer

- 8.4.1 Add 270 mL of ultra-pure water to a clean > 500 mL bottle or other vessel.
- 8.4.2 Add the entire 30 mL contents of the **10X Wash Buffer** bottle to the water.
- 8.4.3 Seal and mix gently by inversion. Avoid foaming or bubbles.
- 8.4.4 Store the **1X Wash Buffer** at room temperature until ready to use in the procedure. Store the prepared **1X Wash Buffer** at 4°C for no longer than 1 week. Do not freeze.

#### 8.5 Microplate Preparation

- Micro-plates are provided ready to use and do not require rinsing or blocking.
- Unused well strips should be returned to the original packaging, sealed and stored at 4°C.
- Equilibrate microplates to ambient temperatures prior to opening to reduce potential condensation.

## 9. Sample Preparation

#### 9.1 Sample Preparation and Storage

- Store samples to be assayed at 2-8°C for 24 hours prior being assayed.
- For long term storage, aliquot and freeze samples at -20°C. Avoid repeated freeze-thaw cycles.
- Clear samples by centrifugation as follows:
  - Cell culture supernatant Remove particulates by centrifugation, assay immediately or aliquot and store samples at -20°C.
  - **Serum** Allow the serum to clot in a serum separator tube (about 4 hours) at room temperature. Centrifuge at approximately 1,000 x g for 15 min. Analyze the serum immediately or aliquot and store samples at -20°C.
  - **Plasma** Collect plasma using heparin, EDTA or Citrate as an anticoagulant. Centrifuge for 15 min at 1,500 x g within 30 min of collection. Assay immediately or aliquot and store samples at -20°C.

#### 9.2 Sample Dilution

Target protein concentration must be estimated and appropriate sample dilution selected such that the final target protein concentration falls near the middle of the assay linear dynamic range.

- Prepare 150 µL sample for each replicate to be assayed.
- Dilute samples with Sample Diluent Buffer.
- Mix diluted samples gently and thoroughly.
- Pipetting less than 2 µL is not recommended for optimal assay accuracy.
- Refer to the following table for recommended sample dilution guidelines based on the dynamic range of this kit:

Estimated Sample To	arget Concentration	Dilution Level	Sample Volume For Two Replicates	Sample Diluent Buffer For Two Replicates	
High Concentration	80-800 ng/mL	1:100	3 µL	297 μL	
Medium Concentration 8-80 ng/mL		1:10	30 µL	270 μL	
Low Concentration	125-8,000 pg/mL	1:2	150 µL	150 μL	
Very Low Concentration	≤125 pg/mL	1:2 or No Dilution	-	-	



## 10. Assay Procedure

- Equilibrate all reagents and materials to ambient room temperature prior to use in the procedure.
- To control for small potential variations in micro well-plate and day to day ambient temperature fluctuations, equilibrate all reagents prior to use and perform all incubation steps at to 37°C for optimal consistency and reproducibility.
- **10.1** Add 100 μL of serially titrated standards, diluted samples or blank into wells of the pre-coated well plate. At least two replicates of each standard, sample or blank is recommended.
- **10.2** Cover the plate with the well plate lid and incubate for 90 minutes.
- **10.3** Remove the plate lid and discard the liquid in the wells by rigorously flicking into an acceptable waste receptacle.
- **10.4** Gently blot any remaining liquid from the wells by tapping inverted on the benchtop onto paper toweling. Do not allow the wells to completely dry at any time.
- 10.5 Add 100 µL of prepared 1X Biotinylated Anti-Mouse Gp130 Antibody to each well.
- **10.6** Cover with the well-plate lid and Incubate for 60 minutes.
- **10.7** Discard the liquid in the wells by rigorously flicking into an acceptable waste receptacle.
- **10.8** Gently blot any remaining liquid from the wells by tapping inverted on the benchtop onto paper toweling. Do not allow the wells to completely dry at any time.
- 10.9 Wash plate 3 times with 1X Wash Buffer as follows:
  - 10.9.1 Add 300  $\mu L$  of **1X Wash Buffer** to each assay well.
  - 10.9.2 Incubate for 1 minute.
  - 10.9.3 Discard the liquid in the wells by rigorously flicking into an acceptable waste receptacle.
  - 10.9.4 Gently blot any remaining liquid from the wells by tapping inverted on the benchtop onto paper toweling. Do not allow the wells to completely dry at any time.
  - 10.9.5 Repeat steps 10.9.1 through 10.9.4 two more times
- **10.10** Add 100 μL of prepared **1X Avidin-Biotin-Peroxidase Complex (ABC)** into each well and incubate for 30 minutes.
- **10.11** Discard the liquid in the wells by rigorously flicking into an acceptable waste receptacle.
- **10.12** Gently blot any remaining liquid from the wells by tapping inverted on the benchtop onto paper toweling. Do not allow the wells to completely dry at any time.
- 10.13 Wash plate 5 times with 1X Wash Buffer as follows:
  - 10.13.1 Add 300 µL of 1X Wash Buffer to each assay well.
  - 10.13.2 Incubate for 1 minute.
  - 10.13.3 Discard the liquid in the wells by rigorously flicking into an acceptable waste receptacle.
  - 10.13.4 Gently blot any remaining liquid from the wells by tapping inverted on the benchtop onto paper toweling. Do not allow the wells to completely dry at any time.
  - 10.13.5 Repeat steps 10.13.1 through 10.13.4 **four** more times.
- **10.14** Add 90 μL of **TMB Color Developing Agent** to each well and incubate in the dark for 25 30 minutes. (NOTE: optimal incubation time must be determined by the user. Optimal development can be visualized by blue shading in the top four standard wells, while the remaining standards are still clear.)
- 10.15 Add 100 µL of TMP Stop Solution to each well. Well color should change to yellow immediately.
- **10.16** Read the O.D. absorbance at 450 nm with a standard microplate reader within 30 minutes of stopping the reaction in step 10.15.



#### 11. Calculation of Results

For analysis of the assay results, calculate the Relative OD<sub>450</sub> for each test or standard well as follows:

(Relative 
$$OD_{450}$$
) = (Well  $OD_{450}$ )-(Mean Blank Well  $OD_{450}$ )

The standard curve is generated by plotting the mean replicate **Relative OD**<sub>450</sub> of each standard serial dilution point vs. the respective standard concentration. The Mouse Gp130 concentration contained in the samples can be interpolated by using linear regression of each mean sample **Relative OD**<sub>450</sub> against the standard curve. This is best achieved using curve fitting software.


**Note:** if the samples measured were diluted, multiply the derived mean sample concentration by the dilution factor for a final sample concentration.

## 12. Typical Expected Data

**12.1 Typical absorbance values**. Expected absorbance for standards when TMB incubation is performed for 20 minutes at 37°C and measured at OD<sub>450</sub>.

Standard Number	8	7	6	5	4	3	2	1
Concentration (pg/mL)	0	125.	250.	500.	1,000.	2,000	4,000	8,000
OD <sub>450</sub>	0.088	0.133	0.199	0.307	0.484	0.829	1.254	1.911

**12.2 Typical standard curve**. This standard curve is for demonstration purposes only. An assay specific standard curve should be performed with each assay.





## 12.3 General Specifications

General Specifications					
Range	125 pg/mL - 8,000 pg/mL				
Sensitivity	< 10 pg/mL (Derived by linear regression of OD <sub>450</sub> of the Mean Blank + 2xSD)				
Specificity	Natural and recombinant Human Gp130 UniProt ID: Q00560				
Cross-Reactivity	No detectable cross-reactivity with other relevant proteins				
Recovery	-				
Linearity	-				

## 12.4 Reproducibility

	Intra-Assay				Inter-Assay	
Sample ID	1	2	3	1	2	3
n =	16	16	16	24	24	24
Mean Measured Concentration (pg/mL)	1.3	3.9	5.6	0.95	4.3	6.1
Standard Deviation (pg/mL)	0.057	0.238	0.291	0.055	0.361	0.397
Consistency (%CV)	4.4	6.1	5.2	5.8	8.4	6.5



#### 13. Technical Resources

#### 13.1 References

- 13.1.1 Hibi, M., Murakami, M., Saito, M., Hirano, T., Taga, T., Kishimoto, T. Molecular cloning and expression of an IL-6 signal transducer, gp130. Cell 63: 1149-1157, 1990.
- 13.1.2 Narazaki, M., Yasukawa, K., Saito, T., Ohsugi, Y., Fukui, H., Koishihara, Y., Yancopoulos, G. D., Taga, T., Kishimoto, T. Soluble forms of the interleukin-6 signal-transducing receptor component gp130 in human serum possessing a potential to inhibit signals through membrane-anchored gp130. Blood 82: 1120-1126, 1993.
- 13.1.3 Yoshida K, Taga T, Saito M, Suematsu S, Kumanogoh A, Tanaka T, Fujiwara H, Hirata M, Yamagami T, Nakahata T, Hirabayashi T, Yoneda Y, Tanaka K, Wang W-Z, Mori C, Shiota K, Yoshida N, Kishimoto T (1996).

#### 13.2 Technical Support

#### USA

Aviva Systems Biology, Corp. 5754 Pacific Center Blvd, Suite 201 San Diego, CA 92121

Phone: 858-552-6979 Toll Free: 888-880-0001 Fax: 858-552-6975

Technical support: techsupport@avivasysbio.com

#### China

Beijing AVIVA Systems Biology 6th Floor, B Building, Kaichi Tower #A-2 Jinfu Road. Daxing Industrial Development Zone Beijing, 102600, CHINA

Phone: (86)10-60214720 Fax: (86)10-60214722

E-mail: support@avivasysbio.com.cn

中国地址:北京大兴工业开发区金辅路甲2号凯驰大厦B座6层(102600)

电话: 010-60214720/21 传真: 010-60214722

产品售前咨询及销售: sales@avivasysbio.com.cn

售后及技术支持: support@avivasysbio.com.cn