

Rat GRO/KC ELISA Kit

Catalog #: OKAG00104

Detection and Quantification of Rat Growth-Regulated Protein α (rGRO/KC) Concentrations in Cell Lysates, Sera and Plasma.

Please read the provided manual as suggested experimental protocols may have changed.

Research Purposes Only. Not Intended for Diagnostic or Clinical Procedures.

CONTENTS

Introduction
Assay Principles4
Assay Format5
Assay Restrictions6
Materials Included6
Additional Materials Required7
Health and Safety Precautions7
Storage Information
Sample Preparation and Storage
Sample Experiment Layout10
Immunoassay Protocol11
Summarized Protocol16
Sensitivity17
Cross Reactivity and Specificity17
Technical Support18
ELISA Plate Template19
Notes20

INTRODUCTION


Rat GRO/KCor Growth-Regulated Protein α , also known as C-X-C Motif Chemokine 1 and Platelet-Derived Growth Factor-Inducible Protein KC, is a 96 amino acid cytokine encoded by the *Cxcl1* gene located at locus 14p22 on chromosome 14. After initial synthesis of the peptide, the 24 residue signal sequence is cleaved, allowing the 72 residue GRO/KC protein to fold properly. GRO/KC, an intercrine alpha (chemokine CxC) family member, essentially has chemotactic activity for neutrophils and contributes to neutrophil activation during inflammation. Moreover, the monomeric and homodimeric structures of GRO/KC are induced by lipopolysaccharides (LPS) and inflammation for expression in the lungs and trachea.

Source: Entrez Gene: GRO/KC epidermal growth factor [Rattus norvegicus]; Swiss-Prot: 14095

ASSAY PRINCIPLES

The Aviva Rat GRO/KC ELISA Kit contains the components necessary for quantitative determination of natural or recombinant rGRO/KC concentrations within any experimental sample including cell lysates, serum and plasma. This particular immunoassay utilizes the quantitative technique of a "Sandwich" Enzyme-Linked Immunosorbent Assay (ELISA) where the target protein (antigen) is bound in a "sandwich" format by the primary capture antibodies coated to each well-bottom and the secondary detection antibodies added subsequently by the investigator. The capture antibodies coated to the bottom of each well are specific for a particular epitope on theRatGRO/KCcytokine while the useradded detection antibodies bind to epitopes on the captured target protein. Amid each step of the procedure, a series of wash steps must be performed to ensure the elimination of non-specific binding between proteins to other proteins or to the solid phase. After incubation and "sandwiching" of the target antigen, a peroxidase enzyme is conjugated to the constant heavy chain of the secondary antibody (either covalently or via Avidin/Streptavidin-Biotin interactions), allowing for a colorimetricreaction to ensue upon substrate addition. When the substrate TMB (3, 3', 5, 5'-Tetramethylbenzidine) is added, the reaction catalyzed by peroxidase yields a blue color that is representative of the antigen concentration. Upon sufficient color development, the reaction can be terminated through addition of Stop Solution (2 N Sulfuric Acid) where the color of the solution will turn yellow. The absorbance of each well can then be read by a spectrophotometer, allowing for generation of a standard curve and subsequent determination of protein concentration.


ASSAY FORMAT






Blue TMB Diimine Product







 $\dot{}$


Monoclonal capture antibodies specific for the target are adsorbed to the plate.

Target antigen present in standard or sample is bound by capture antibodies on the solidphase.

Polyclonal, biotinlabeled detection antibodies specific for the target are added to bind another epitope on the target antigen.

Avidin-HRP attaches to detection antibody via high affinity avidinbiotin interaction.

TMB substrate is converted to the blue TMB diimine via the HRP enzyme. Upon addition of acid, the reaction terminates and the wells can be read at 450 nm.



ASSAY RESTRICTIONS

- This ELISA kit is intended for research purposes only, NOT diagnostic or clinical procedures of any kind.
- Materials included in this kit should NOT be used past the expiration date on the kit label.
- Reagents or substrates included in this kit should NOT be mixed or substituted with reagents or substrates from any other kits.
- Variations in pipetting technique, washing technique, operator laboratory technique, kit age, incubation time or temperature may cause differences in binding affinity of the materials provided.
- The assay is designed to eliminate interference and background by other cellular macromolecules or factors present within any biological samples. However, the possibility of background noise cannot be fully excluded until all factors have been tested using the assay kit.

Component	Quantity Per Plate	Container -	
Microstrips Coated w/ Capture Antibody	12 x 8-well MicroStrips		
Protein Standard	Lyophilized (83 ng)	Red	
Biotinylated Detection Antibody	Lyophilized	Yellow	
400X Streptavidin-HRP	30 µl	Blue	
Wash Buffer (15x)	50 ml	Clear	
Assay Diluent	50 ml	Clear	
Ready-to-Use Substrate	12 ml	Brown	
Stop Solution	12 ml	Clear	
Adhesive Plate Sealers	4 Sheets	-	
Technical Manual	1 Manual	-	

MATERIALS INCLUDED

ADDITIONAL MATERIALS REQUIRED

The following materials and/or equipment are NOT provided in this kit but are necessary to successfully conduct the experiment:

- Microplate reader able to measure absorbance at 450 nm (with correction wavelength set to 540 nm or 570 nm)
- Micropipettes with capability of measuring volumes ranging from 1 μI to 1 ml
- Deionized or sterile water
- Squirt bottle, manifold dispenser, multichannel pipette reservoir or automated microplate washer
- Graph paper or computer software capable of generating or displaying logarithmic functions
- Absorbent paper or vacuum aspirator
- Test tubes or microfuge tubes capable of storing ≥1 ml
- Bench-top centrifuge (optional)
- Bench-top vortex (optional)
- Orbital shaker (optional)

HEALTH AND SAFETY PRECAUTIONS

- Reagents provided in this kit may be harmful if ingested, inhaled or absorbed through the skin. Please carefully review the MSDS for each reagent before conducting the experiment.
- Stop Solution contains 2 N Sulfuric Acid (H₂SO₄) and is an extremely corrosive agent. Please wear proper eye, hand and face protection when handling this material. When the experiment is finished, be sure to rinse the plate with copious amounts of running water to dilute the Stop Solution prior to disposing the plate.

STORAGE INFORMATION

Note: If used frequently, reagents may be stored at 4°C.

Unopened Kits: Store at 4°C for 6 months.

Component	Storage Time	Storage Information		
Microstrips Coated w/ Capture Antibody				
400X Streptavidin-HRP		4°C		
Wash Buffer (15x)	6 Months			
Assay Diluent				
Ready-to-Use Substrate				
Stop Solution				
Protein Standard	Lyophilized: 6 Months			
Biotinylated Detection Antibody	Reconstituted: 1 Month	4°C		
Adhesive Plate Sealers	-	-		
Technical Manual	-	-		

SAMPLE PREPARATION AND STORAGE

If samples are to be used within 24 hours, aliquot and store at 4°C. If samples are to be used over a long period of time, aliquot and store between -20°C and -80°C, depending on the duration of storage.

Note: Samples containing a visible precipitate or pellet must be clarified prior to use in the assay.

Caution: Avoid repeated freeze/thaw cycles to prevent loss of biological activity of proteins in experimental samples.

Cell Lysate and Supernatants

Remove large cell components via centrifugation and perform the assay. Cell lysates and supernatants require a dilution using Assay Diluent. A serial dilution may be performed to determine a suitable dilution factor for the sample. For future use of the sample, follow the sample storage guidelines stated above.

Serum

Allow samples to clot in a serum separator tube (SST) for 30 minutes. After sufficient clotting, centrifuge at 1000 x g for 15 minutes and remove serum from SST in preparation for the assay. Serum samples require at least a 1:50 dilution using Assay Diluent. For future use of the sample, follow the storage guidelines above.

Plasma

Use heparin, citrate or EDTA as an anticoagulant to gather plasma from original biological sample. After collection of the plasma, centrifuge for 15 minutes at 1000 x g. This step must be performed within 30 minutes of plasma collection. Plasma samples require at least a 1:50 dilution using Assay Diluent. Afterwards, perform the assay or for future use of the sample, follow the storage guidelines stated above.

SAMPLE EXPERIMENT LAYOUT

	1	2	3	4	5	6
Α	Standard (High Point)	Standard (High Point)	Standard (High Point)	Sample	Sample	Sample
В	Standard (1:2)	Standard (1:2)	Standard (1:2)	Sample	Sample	Sample
С	Standard (1:4)	Standard (1:4)	Standard (1:4)	Sample	Sample	Sample
D	Standard (1:8)	Standard (1:8)	Standard (1:8)	Sample	Sample	Sample
Ε	Standard (1:16)	Standard (1:16)	Standard (1:16)	Sample	Sample	Sample
F	Standard (1:32)	Standard (1:32)	Standard (1:32)	Sample	Sample	Sample
G	Standard (1:64)	Standard (1:64)	Standard (1:64)	Sample	Sample	Sample
Н	Negative Control	Negative Control	Negative Control	Sample	Sample	Sample

IMMUNOASSAY PROTOCOL

Note: If possible, all incubation steps should be performed on an orbital shaker to equilibrate solutions when added to the microplate wells. Also, all provided solutions should be at ambient temperature prior to use.

Note:Avoid adding solutions into wells at an angle, always keep pipette tip perpendicular to plate bottom.

Reconstitution of Provided Materials


- 1. Reconstitute the Biotin-Conjugated Detection Antibody in 100 μ l of ddH₂O for a concentration of 83 μ g/ml.
- 2. Reconstitute the Protein Standard in 83 μ l of ddH₂O for a concentration of 1 μ g/ml.
- 3. Dilute the 50 ml of 15X Wash Buffer in 700 ml of ddH2O for 750ml of 1X Wash Buffer.

Addition of Known Standard and Unknown Sample to Immunoassay


The Rat GRO/KC ELISA Kit allows for the detection and quantification of endogenous levels of natural and/or recombinant Rat GRO/KC proteins within the range of 16-1000 pg/ml.

1. Dilute the known standard sample from 1000 pg/ml to 0 pg/ml in a series of microfuge tubes. Mix each tube thoroughly by inverting several times or by vortexing lightly to ensure proper equilibration. Add 100 µl of each serial dilution step into the wells of a specified row or column of the 96-well microtiter plate in duplicate or triplicate and incubate at room temperature for 2 hours. Unknown samples of interest can be serial diluted with Assay Diluent to concentrations within the detection range of this assay kit and added to the plate at 100 µl per well. Seal the microplate air-tight using one of the microplate adhesive seals provided in this kit or Parafilm if readily available. See Appendix for serial dilution diagram.

To obtain serial dilution high point, dilute reconstituted Protein Standard to the maximum concentration for serial dilution by adding $n \mu l$ reconstituted Protein Standard to serial dilution high point tube and then raising the volume to 200 μ l. Shown below is a diagram illustrating a hypothetical 2-fold serial dilution on a given reconstituted Protein Standard.



For samples of unknown protein concentrations, serial dilute the experimental sample using Assay Diluent to determine range of detection and acceptable dilutions. Shown below is a diagram illustrating a 10-fold serial dilution on a given Sample of Interest.



Addition of Detection Antibody to Capture Antibody-Bound Samples

- Aspirate the protein standard solution out of the microplate wells. If your lab does not have a vacuum-based aspirator, you may dump the solutions from the microplate into a waste container and blot 3-4 times on a stack of paper towels until most or all of the liquid is removed from the wells. Dilute the 15X wash buffer to 1Xusing pure H₂O. Add 300-400 µl of 1X Wash Buffer to each well being used and gently shake for 5-7 minutes on an orbital shaker. Perform this wash step 4 times consecutively.
- 2. After the 4th wash step, dilute the detection antibody solution 1:100 in Assay Diluent to a concentration of 1 μ g/ml. Mix the test tube either by inverting several times or vortexing to ensure proper equilibration. Ensure that there is enough detection antibody solution for all wells being used. Add 100 μ l of the diluted detection antibody solution into each well, seal the plate and incubate at room temperature for 2 hours.

Conjugation of Streptavidin-HRP to Biotinylated Detection Antibody

- 1. Remove the detection antibody solution out of the microplate wells by either vacuum-based aspirator or paper towel blotting. Perform 4 consecutive wash steps with gentle shaking between each wash.
- 2. Dilute the 400x Streptavidin-HRP by 1:400 using Assay Diluent to a 1x Streptavidin-HRP Solution.
- 3. After the 4th wash step, add 100 µl of 1x Streptavidin-HRP Solution into each well and incubate at room temperature for 30 minutes.

Application of Liquid Substrate for Colorimetric Reaction

1. Remove the 1X Streptavidin-HRP solution out of the microplate wells by either vacuum-based aspirator or paper towel blotting. Prepare the Ready-to-Use substrate by bringing it to room temperature without exposure to fluorescent or UV light as these may degrade the substrate. Perform 4 consecutive wash steps with gentle shaking between each wash.


- 2. After the 4th wash step, add 100 µl of Ready-to-Use Substrate solution into each well and incubate at room temperature for color development. The microplate should be kept out of direct light by either covering with an opaque object or putting it into a dark room. Closely monitor the color development as some wells may turn blue very quickly depending on analyte and/or detection antibody-HRP concentrations. Once the blue color has ceased to develop further, immediately add 100 µl of Stop Solution to each well being used. The color in the wells should immediately change from blue to yellow.
- 3. The microplate is now ready to be read by a microplate reader. Within 30 minutes of adding the Stop Solution, determine the optical density (absorbance) of each well by reading the plate with the microplate reader set to 450 nm. If wavelength correction is available, set to 540 nm or 570 nm. If wavelength correction is not available, subtract readings at 540 nm or 570 nm from the readings at 450 nm. **Caution:** Readings made directly at 450 nm without correction may be higher and less accurate.

Generation of Standard Curve and Interpretation of Data

- 1. Average the duplicate or triplicate readings for each standard, control and sample and subtract the average zero standard optical density.
- 2. Generate a standard curve by using Microsoft Excel or other computer software capable of establishing a 4-Parameter Logistic (4-PL) curve fit. If using Excel or an alternative graphing tool, plot the average optical density values in absorbance units (y-axis) against the known standard concentrations in pg/ml (x-axis). Note: Only use the values in which a noticeable gradient can be established. Afterwards, generate a best fit curve or "trend-line" through the plotted points via regression analysis. Note: Shown on the next page is an example of typical data produced by analysis of the standard sample.

The data and subsequent graph was obtained after performing a cytokine ELISA for RatGRO/KC.Each known sample concentration was assayed in triplicate.

RatGRO/KCStandard Curve					
Concentration (pg/ml)	Average OD 450nm				
1000	1.6				
500	1.1				
250	0.7				
125	0.42				
62.5	0.25				
31.25	0.15				



SUMMARIZED PROTOCOL

Reconstitute Biotinylated Detection Antibody and Protein Standard and dilute the 10X Wash Buffer as specified.

Perform serial dilution of Protein Standard and prepare samples as desired. See sample preparation section for instructions to dilute serum and plasma samples.

Add 100 µl of Protein Standard, sample or controlto each well and incubate for 2 hours at room temperature.

Aspirate Protein Standards, samples or controls out and wash plate 4 times.

Dilute Biotinylated Detection Antibody as specified. Add 100 µl to each well and incubate for 2 hours at room temperature.

Aspirate Biotinylated Detection Antibody out and wash plate 4 times.

Dilute 400x Streptavidin-HRP as specified. Add 100 µl of 1X Streptavidin-HRP to each well and incubate at room temperature for 30 minutes.

Aspirate 1X Streptavidin-HRP out and wash plate 4 times.

Add 100 µl of Ready-to-Use Substrate to each well and incubate at room temperature for color development.



Add 100 µl of Stop Solution and read plate at 450 nm.











SENSITIVITY

The Rat GRO/KC ELISA Kit allows for the detection and quantification of endogenous levels of natural and/or recombinant Rat GRO/KC proteins within the range of 16-1000 pg/ml.

CROSS REACTIVITY AND SPECIFICITY

The RatGRO/KCELISA is capable of recognizing both recombinant and naturally produced RatGRO/KCproteins. The antigens listed below were tested at 50 ng/ml and exhibited 100% cross reactivity.

• Murine:KC

The antigens listed below were tested at 50 ng/ml and exhibited less than 1% cross reactivity or interference.

- Human: GRO/MGSA, GROβ, GROγ
- Murine: MIP-2
- **Rat**: GROβ/MIP-2

TECHNICAL SUPPORT

For troubleshooting, information or assistance, please go online to <u>www.avivasysbio.com</u> or contact us at:

Aviva Systems Biology Corporation 5754 Pacific Center Blvd., Suite 201 San Diego, CA 92121 United States of America

Email: info@avivasysbio.com or techsupport@avivasysbio.com

Phone: (858) 552-6979 Toll-Free Phone: (888) 880-0001 Fax: (858) 552-6975

ELISA PLATE TEMPLATE

Ŧ	G	Ψ	m	D	C		A	
								-
	9					3 5		2
								ω
								4
								u.
								σ
								-
								8
								v
	<u>.</u>							5
						<u>.</u>		H
	N					A		5

NOTES

| Rat GRO/KC| 20