



## Certificate of Analysis

---

Product: CA 19-9 EIA TEST KIT  
Catalog Number: OKBA00005  
Lot Number: KG0007  
Expiration Date: 06-2023

Kit Component	Lot No.	Contents	Exp. Date
CA 19-9 Ab Coated Wells	KG0007	1 ea	06-2023
Standard Set	KG0007	1 set	06-2023
Conjugate Conc. (12X)	KG0007	1.1 mL	06-2023
Conjugate Diluent	KG0007	13 mL	06-2023
Assay Buffer	KG0007	13 mL	06-2023
TMB Reagent	KG0007	11 mL	06-2023
Stop Solution (in HCl)	KG0007	11 mL	06-2023
Wash Buffer (20X)	KG0007	50 mL	06-2023
Bio-Rad Controls (not included)	---	---	---

### Standard Curve

Concentration (U/mL)	O.D. 450 nm
0	0.050
25	0.265
75	0.674
150	1.205
300	1.922
600	2.850

This standard Curve was obtained on the date of performance testing of the above components. It is for the purpose of illustration only. Each laboratory should establish its own standard curve.

### Control Values

Control No.	Mean (U/mL)	Range
1	11.4	3.64 – 18.4
2	40.2	28.2 – 50.8
3	97.4	78.1 – 117

---

### AVIVA SYSTEMS BIOLOGY

10211 Pacific Mesa Blvd, Suite 401 San Diego, CA 92121 USA | Tel: (858)552-6979 | [techsupport@avivasysbio.com](mailto:techsupport@avivasysbio.com)