



Certificate of Analysis

Product: CA125 EIA TEST KIT
Catalog Number: OKBA00006
Lot Number: KF1156
Expiration Date: 10-2022

Kit Component	Lot No.	Contents	Exp. Date
CA125 Ab Coated Wells	KF1156	1 ea	10-2022
Standards, Liq.	KF1156	1 set	10-2022
Conjugate Reagent	KF1156	13 ml	10-2022
TMB Reagent	KF1156	11 ml	10-2022
Stop Solution	KF1156	11 ml	10-2022
20X Wash Buffer	KF1156	50 ml	10-2022
Bio-Rad Controls (not included)	---	---	---

Standard Curve

Concentration (U/ml)	O.D. 450 nm
0	0.060
15	0.141
50	0.410
100	0.818
200	1.498
400	2.576

This standard Curve was obtained on the date of performance testing of the above components. It is for the purpose of illustration only. Each laboratory should establish its own standard curve.

Control Values

Control No.	Mean (U/ml)	Range
1	29.6	17.0 – 35.1
2	69.5	45.8 – 85.3
3	200	135 – 259