



Certificate of Analysis

Product: Free Beta-hCG EIA TEST KIT
Catalog Number: OKBA00009
Lot Number: KG0046
Expiration Date: 02-2023

Kit Component	Lot No.	Contents	Exp. Date
Beta-hCG Ab Coated Wells	KG0046	1 ea	02-2023
Standards, Lyo.	KG0046	1 set	02-2023
Conjugate Reagent	KG0046	18 ml	02-2023
TMB Reagent	KG0046	11 ml	02-2023
Stop Solution	KG0046	11 ml	02-2023
Zero Buffer	KG0046	13 ml	02-2023
Bio-Rad Controls (not included)	---	---	---

Standard Curve

Concentration (ng/ml)	O.D. 450 nm
0	0.052
2.5	0.311
5.0	0.551
10.0	0.951
25.0	2.015
50.0	3.062

This standard Curve was obtained on the date of performance testing of the above components. It is for the purpose of illustration only. Each laboratory should establish its own standard curve.

Control Values

Control No.	Mean (ng/ml)	Range
1	4.52	2.73 – 5.10
2	15.4	10.3 – 18.6
3	33.2	21.7 – 40.3