



## Certificate of Analysis

---

Product: FSH EIA TEST KIT  
Catalog Number: OKBA00011  
Lot Number: KF0071  
Expiration Date: 02-2022

Kit Component	Lot No.	Contents	Exp. Date
FSH Ab Coated Wells	KF0071	1 ea	02-2022
Standards, Lyo.	KF0071	1 set	02-2022
Conjugate Reagent	KF0071	13 ml	02-2022
TMB Reagent	KF0071	11 ml	02-2022
Stop Solution	KF0071	11 ml	02-2022
Bio-Rad Controls (not included)	---	---	---

### Standard Curve

Concentration (mIU/ml)	O.D. 450 nm
0	0.049
5	0.096
15	0.219
50	0.859
100	1.566
200	2.959

This standard Curve was obtained on the date of performance testing of the above components. It is for the purpose of illustration only. Each laboratory should establish its own standard curve.

### Control Values

Control No.	Mean (mIU/ml)	Range
1	12.1	6.54 – 14.6
2	25.2	18.5 – 34.6
3	68.5	45.4 – 89.0

---

**AVIVA SYSTEMS BIOLOGY**

7700 Ronson Rd., Suite 100 San Diego, CA 92111 USA | Tel: (858)552-6979 | techsupport@avivasysbio.com