



Certificate of Analysis

Product: LH EIA TEST KIT
Catalog Number: OKBA00018
Lot Number: KF0111
Expiration Date: 01-2022

Kit Component	Lot No.	Contents	Exp. Date
LH Ab Coated Wells	KF0111	1 ea	01-2022
Standards, Lyo.	KF0111	1 set	01-2022
Conjugate Reagent	KF0111	13 ml	01-2022
TMB Reagent	KF0111	11 ml	01-2022
Stop Solution	KF0111	11 ml	01-2022
Bio-Rad Controls (not included)	370	---	---

Standard Curve

Concentration (mIU/ml)	O.D. 450 nm
0	0.050
5	0.139
15	0.354
50	0.966
100	1.831
200	2.982

This standard Curve was obtained on the date of performance testing of the above components. It is for the purpose of illustration only. Each laboratory should establish its own standard curve.

Control Values

Control No.	Mean (mIU/ml)	Range
1	3.04	1.31 – 4.10
2	20.7	13.3 – 33.2
3	65.4	42.3 – 95.1

AVIVA SYSTEMS BIOLOGY

7700 Ronson Rd, Suite 100 San Diego, CA 92111 USA | Tel: (858)552-6979 | info@avivasysbio.com