



Certificate of Analysis

Product: T4 EIA TEST KIT
Catalog Number: OKBA00024
Lot Number: KF0073
Expiration Date: 07-2022

Kit Component	Lot No.	Contents	Exp. Date
T4 Ab Coated Wells	KF0073	1 ea	07-2022
Standard Set	KF0073	1 set	07-2022
Conjugate Conc. (11X)	KF0073	1.3 mL	07-2022
Conjugate Reagent	KF0073	13 mL	07-2022
TMB Reagent	KF0073	11 mL	07-2022
Stop Solution	KF0073	11 mL	07-2022
Bio-Rad Controls (not included)	---	---	---

Standard Curve

Concentration ($\mu\text{g/dL}$)	O.D. 450 nm
0	2.719
2	1.723
5	1.222
10	0.905
15	0.689
25	0.354

This standard Curve was obtained on the date of performance testing of the above components. It is for the purpose of illustration only. Each laboratory should establish its own standard curve.

Control Values

Control No.	Mean ($\mu\text{g/dL}$)	Range
1	4.00	2.37 – 8.38
2	9.09	5.59 – 14.9
3	8.59	6.94 – 15.8

AVIVA SYSTEMS BIOLOGY

7700 Ronson Rd, Suite 100 San Diego, CA 92111 USA | Tel: (858)552-6979 | techsupport@avivasysbio.com