

Certificate of Analysis

Product Name	Human enterovirus 71 virus (EV71) antibody (IgG) ELISA kit		
Catalog	OKCA00455		
Lot	KF0455		
Test parameter	specification	test result	approval
Positive control	>0.6	1.574	approved
Negative control	<0.15	0.073	approved
CV	$\leq 15\%$	10.59%	approved
Positive rate	10, Positive	100%	approved
Negative rate	10, Negative	100%	approved