

## Certificate of Analysis

Product Name	Human salmonella typhi (ST) antibody (IgM) ELISA kit		
Catalog	OKCA00520		
Lot	KE1433		
Test parameter	specification	test result	approval
Positive control	>0.4	0.954	approved
Negative control	<0.21	0.082	approved
CV	≤15%	10.56%	approved
Positive rate	10, Positive	100%	approved
Negative rate	10, Negative	100%	approved