



## CERTIFICATE OF ANALYSIS

**Product:** Mouse BDNF ELISA Kit

**Catalog No:** OKEH00022

**Lot No:** KG0130

**Expiration Date:** 2022.09

**Performance Characteristics:**

Intra-Assay CV	< 5.2%	Passed
Inter-Assay CV	< 8.4%	Passed
Limit of Detection (LOD)		< 0.08 ng/mL
Assay Range		0.156 – 10 ng/mL

**Standard Dilution:**


Prepare the standards immediately before use. For each experiment, add 2 mL of Sample Diluent to the 40 ng Lyophilized BDNF Standard. Mix gently until dissolved. Leave the vial at ambient temperature for 15 minutes.

Prepare a 10 ng/mL Standard by adding 500 µL of 20 ng/mL Standard to 500 µL of Sample Diluent in Tube #1. Mix gently and thoroughly.

Standard Number (Tube)	Standard to Dilute	Volume Standard to Dilute (µL)	Volume Sample Diluent (µL)	Total Volume (µL)	Final Concentration (ng/mL)
Standard	40 ng BDNF Standard	NA	2,000	2,000	20 ng/mL
1	20 ng/mL	500	500	1,000	10 ng/mL
2	10 ng/mL	300	300	600	5.0 ng/mL
3	5.0 ng/mL	300	300	600	2.5 ng/mL
4	2.5 ng/mL	300	300	600	1.25 ng/mL
5	1.25 ng/mL	300	300	600	0.625 ng/mL
6	0.625 ng/mL	300	300	600	0.312 ng/mL
7	0.312 ng/mL	300	300	600	0.156 ng/mL
8	N/A	0	300	300	0.0 (Blank)

Sample Standard Curve and Calculations ONLY. Do not use below report for result calculations. A standard curve should be generated each time the test is performed

### Standard Curve Graph



Standard ng/mL	Absorbance (OD 450nm)
10	2.752
5.0	2.148
2.5	1.248
1.25	0.728
0.625	0.424
0.312	0.196
0.156	0.101