

CERTIFICATE OF ANALYSIS

Product: S-Adenosyl methionine ELISA Kit (General)
SKU: OKEH02593
Lot: KG0375
Expiration Date: 2022.12

Performance Characteristics:

Intra-Assay CV	<= 5.4%	Passed
Inter-Assay CV	<= 8.1%	Passed
Limit of Detection (LOD)		<= 0.38 nmol/L
Assay Range		1.56 - 100 nmol


Standard Preparation:

Prepare the standards immediately before use. For each experiment, add 1 mL of Sample Diluent to the 100 nmol Lyophilized S-Adenosyl methionine standard. Mix gently until dissolved. Leave the vial at ambient temperature for 15 minutes.

Prepare a set of diluted standards as indicated in the chart below:

Standard Number (Tube)	Standard To Dilute	Volume Standard To Dilute (uL)	Volume Sample Diluent Buffer (uL)	Total Volume (uL)	Final Concentration (nmol/L)
1	Reconstitute S-Adenosyl methionine Standard	N/A	N/A	1,000	100

Standard Number (Tube)	Standard To Dilute	Volume Standard To Dilute (uL)	Volume Sample Diluent Buffer (uL)	Total Volume (uL)	Final Concentration (nmol/L)
2	100	300	300	600	50
3	50	300	300	600	25
4	25	300	300	600	12.5
5	12.5	300	300	600	6.25
6	6.25	300	300	600	3.12
7	3.12	300	300	600	1.56
8	N/A	0	300	300	0.0 (Blank)



Typical Expected Data

Standard Curve and Calculations ONLY. Do not use below report for result calculations. A standard curve should be generated each time the test is performed.

Standard Number	Concentration (nmol/L)	Background Adjusted OD Value
1	100	0.25
2	50	0.342
3	25	0.438
4	12.5	0.607
5	6.25	0.889

Standard Number	Concentration (nmol/L)	Background Adjusted OD Value
6	3.12	1.361
7	1.56	1.972

Standard Curve Graph

