

CERTIFICATE OF ANALYSIS

Product: UTS2 ELISA Kit (Mouse)
SKU: OKEH05846
Lot: BJ03192021
Expiration Date: 2021.09

Performance Characteristics:

Intra-Assay CV	<= 4.6%	Passed
Inter-Assay CV	<= 7.8%	Passed
Limit of Detection (LOD)		<= 15.75 pg/mL
Assay Range		31.2 - 2000 pg/mL

Standard Preparation:

Prepare the standards immediately before use. For each experiment, add 1 mL of Sample Diluent to the 2000 pg Lyophilized UTS2 standard. Mix gently until dissolved. Leave the vial at ambient temperature for 15 minutes.

Prepare a set of diluted standards as indicated in the chart below:

Standard Number (Tube)	Standard To Dilute	Volume Standard To Dilute (uL)	Volume Sample Diluent Buffer (uL)	Total Volume (uL)	Final Concentration (pg/mL)
1	Reconstitute UTS2 Standard	N/A	N/A	1,000	2000
2	2000	300	300	600	1000

Standard Number (Tube)	Standard To Dilute	Volume Standard To Dilute (uL)	Volume Sample Diluent Buffer (uL)	Total Volume (uL)	Final Concentration (pg/mL)
3	1000	300	300	600	500
4	500	300	300	600	250
5	250	300	300	600	125
6	125	300	300	600	62.5
7	62.5	300	300	600	31.2
8	N/A	0	300	300	0.0 (Blank)



Typical Expected Data

Standard Curve and Calculations ONLY. Do not use below report for result calculations. A standard curve should be generated each time the test is performed.

Standard Number	Concentration (pg/mL)	Background Adjusted OD Value
1	2000	2.667
2	1000	1.791
3	500	0.938
4	250	0.528
5	125	0.269
6	62.5	0.128

Standard Number	Concentration (pg/mL)	Background Adjusted OD Value
7	31.2	0.063

Standard Curve Graph

